

State of California
AIR RESOURCES BOARD

EXECUTIVE ORDER I-08-028

Relating to Independent Contractor Approval under
Section 91207, Title 17, California Code of Regulations

Datalab

WHEREAS, the Air Resources Board (ARB), pursuant to section 41512 of the California Health and Safety Code, has established the procedures contained in sections 91200-91220, title 17, California Code of Regulations, to allow the use of independent testers for compliance tests required by ARB; and

WHEREAS, it has been determined that Datalab meets the requirements of ARB for conducting ARB Test Methods 1, and 4 pursuant to sections 91200-91220, title 17, California Code of Regulations when the following conditions are met:

1. Datalab calibrates its temperature gauges in accordance with section 4.3 of ARB Test Method 2 and establishes and maintains a log of the calibrations;
2. Datalab calibrates its dry gas meters as required by section 5.3 of ARB Test Method 5, and establishes and maintains a log of the calibrations;
3. Datalab measures barometer pressure as described in section 2.5 of ARB Test Method 2. It must calibrate its barometer in accordance with section 4.4 of ARB Test Method 2 and establishes and maintains a log of the calibrations;

WHEREAS, ARB's Executive Officer, pursuant to Health and Safety Code section 39516, issued Executive Order G-02-008, delegating to the Chief of ARB's Monitoring and Laboratory Division the authority to approve independent testers in accordance with title 17, California Code of Regulations, sections 91200-91220.

NOW, THEREFORE, I, William V. Loscutoff, Chief of ARB's Monitoring and Laboratory Division, order that Datalab is granted an approval, from the date of execution of this order, until June 30, 2009, to conduct the tests listed above, subject to compliance with sections 91200-91220, title 17, California Code of Regulations.

BE IT FURTHER ORDERED that during the approved period the Executive Officer or his authorized representative may field audit one or more tests conducted pursuant to this order for each type of testing listed above.

Executed at Sacramento, California this 8th day of July 2008.

Original signed by

William V. Loscutoff, Chief
Monitoring and Laboratory Division